

Parallel Guide KIT

to correct location and direction of Implant

OSSTEM[®]
IMPLANT

Prime KIT: OPGPK (options excluded)

Advanced KIT: OPGAk (including all options)



Guide Drill



Bridge Guide



Parallel Guide KIT



Single Guide



Guide Pin

Composition of Lower Valve



L-Wrench Option



Denture Guide Option



Multi Joint Handle Option

Guide Drill

- Initial drill : fastened to the single guide for initial drilling, and enables drilling to be stopped at a depth of 7 mm
- Twist drill : used with bridge guide for initial drilling, and enables to the depth to be adjusted by referencing marking

D	Ø2.2
Initial drill	GD2208NC
Twist drill	GS2213FNC



Initial drill



Twist drill

Single Guide

- Transparent material used to recognize location and direction of drilling
- Function of initial drill stops at 7 mm depth
- 6-type composition from Ø5.0 to 10.0 based on mesio-distal crown diameter

D	Ø5.0	Ø6.0	Ø7.0	Ø8.0	Ø9.0	Ø10.0
	SGB050	SGB060	SGB070	SGB080	SGB090	SGB100



Ø5.0



Ø6.0



Ø7.0



Ø8.0



Ø9.0



Ø10.0

Guide Pin

- Checking the path on the drilling site and fixing the single guide

PGSP22




Bridge Guide

- Adjusting distance and direction of drilling
- Fan type: selecting from 7.0 - 12.5 mm interval, in increments of 0.5 mm
- Compass type: adjustable from 5 - 24 mm interval, in increments of 1 mm
- Use after controlling distance in the distance setup at the middle plate of KIT



Type \ Distance	7~8.5	9~10.5	11~12.5	5~24
Fan	PGBRA070	PGBRA090	PGBRA110	-
Compass	-	-	-	PGBPA



Bridge guide is fan-shaped, enabling easy adjustment of distance

Multi Joint Handle Option

- Equipment to hold the location of guide from out of mouth by being connected to the ball head of the Bridge guide

MJH



Denture Guide Option

- Angle of guide is adjustable for each individual patient in cases of edentulous jaw
- Oral drilling while fixing the angle with the L-wrench on the plaster cast according to the dental arch shape of patient

PGODA



L-wrench Option

- Equipment to firmly fix the Denture guide after adjusting the size

LWC20H



Single Guide enables correct measurement through close attachment

OSSTEM[®]
IMPLANT

DEUTSCHE OSSTEM GmbH

Mergenthalerallee 25, 65760 Eschborn, Germany

Tel: +49(0)6196 777 550 Fax: +49(0)6196 777 5529

E-mail: info@osstem.de www.osstem.de